

Certificate of Calibration



Glenammer
Laboratory Test Sieves

Email: sales@glenammer.com

Web Site: www.glenammer.com

Glenammer Sieves Ltd
62 Viewfield Road
Ayr
KA8 8HH
Tel: +44 (0)1292 261444



24551

Customer:	Sample	Issued by:	Daniel Smith
Address:	Sample	Date of Issue:	30/08/2023
	Sample	Certificate No.:	CC-0157
	Sample	Calibration Technician:	Daniel Smith
	Sample	Date of Test:	30/08/2023
	Sample	Test Temperature:	20.0 °C ± 2.0 °C
Sieve:	300 Dia. x 25.00mm p/p Sieve	Test Type:	Optical
Condition:	New	Test Method:	Crosswise Spot Check
Serial No.:	23081480	Expanded Uncertainty:	0.004 mm
Specification:	BS ISO 3310-2:2013	Approved Signatory:	Daniel Smith
Lab Location:	Glenammer (see above)	Signature:	

Results

Parameters	Measured Values ⁽¹⁾		Standard Tolerances ⁽²⁾	Decision
	Warp	Weft		
Maximum size ⁽³⁾	25.022 mm	25.010 mm	≤ w ^{a (8)} + 0.350 mm	Accepted
Minimum size ⁽⁴⁾	24.967 mm	24.974 mm	≥ w ^{a (8)} - 0.350 mm	Accepted
Measured holes ⁽⁵⁾	35	35	≥ All or 50	Accepted
Maximum pitch ⁽⁶⁾	31.536 mm	31.510 mm	≤ 36.000 mm	Accepted
Minimum pitch ⁽⁷⁾	31.411 mm	31.476 mm	≥ 28.500 mm	Accepted

(1) Values are measured in two orthogonal directions labelled warp and weft.

(2) Tolerances according to BS ISO 3310-2:2013.

(3) Maximum value measured for hole size.

(4) Minimum value measured for hole size.

(5) Total number of holes measured.

(6) Maximum value measured for pitch.

(7) Minimum value measured for pitch.

(8) Nominal hole size.

Decision Rule:

To account for measurement uncertainty, we constrain the measured values that can be considered acceptable to only values which fall within the acceptance interval for each measurement. The difference between the acceptance interval and the standard tolerances provided by BS ISO 3310-2:2013 is equal to the expanded uncertainty of measurement reported on this certificate. All measured values in this certificate have a single sided acceptance interval. The number of measured holes does not have an associated measurement uncertainty. It is evaluated using a single sided acceptance interval which is equal to the standard tolerances. If all results are accepted, then the sieve is accepted as conforming to the specification.

Based on the above results the sieve identified above is **Accepted** as conforming to BS ISO 3310-2:2013

Calibration Methods:

The calibration of this sieve has been carried out in accordance with the procedures documented in BS ISO 3310-2:2013 using the Optical method. The sieve was sampled using the Crosswise Spot Check method. All test equipment used in this calibration is calibrated and traceable to a UKAS accredited laboratory.

Uncertainty of Measurement:

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$ which for a normal distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

The results shown only relate to the item identified on this certificate.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

End of Report.